

Test Report

REPORT NO: UB/2018/10172

Date: 2018/02/14

Page : 1 of 3



Green Health Canada Inc.

1-1630 Matheson Blvd, Mississauga Ontario, L4W 1Y4 Canada

The following sample(s) was/were submitted and identified by/on behalf of applicant as:

Sample Name: EXA-ONE Catalogue number:GHC 7121

Applicant : Green Health Canada Inc.

Item/Lot No: 171215

Sample Condition: Please refer to the photo for sample shown at the last page of this report.

MFG: ---

EXP: 11-2022

Sample Received: 2018/01/25

Date of Testing: 2018/01/25 to 2018/02/14

Test Results: -Please refer to next page(s)-

Signed for and on behalf of

SGS Taiwan Ltd.

Shin-Jyh Chen

Shin-Jyh Chen

Manager



This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

Test Report

REPORT NO: UB/2018/10172

Date: 2018/02/14

Page : 2 of 3



Green Health Canada Inc.

1-1630 Matheson Blvd, Mississauga Ontario, L4W 1Y4 Canada

Test Results:

Test Item	Test Method	Test Results
Flammability (as Received)	With reference to CPSC 16 CFR 1610-2008 method.	IBE

Test Item	Test Method	Test Results	Unit
Water Resistance Hydrostatic Pressure	With reference to AATCC 127-2013 Option 2 method.(The rate of water pressure:60±3cmH ₂ O/min)	722	mmH ₂ O
Water Resistance Impact Penetration Test	With reference to AATCC 42-2013 Type II method. (Environmental condition:16°C,65%RH)	0	g
Seam Strength	With reference to JIS L1093:2011 7.1.1 A-1 method. (Speed:30±1cm/min,CRE)	6.7	kgf

Test Item	Test Method	Test Results		Unit
		Length wise	Width wise	
Tearing Strength	With reference to ASTM D5733-1999 Option 2 method. (Trapezoid Method Speed:30± 1cm/min,CRE)	3.6	1.9	kgf
Breaking Strength	With reference to ASTM D5034-2009 method. (Grab Method Speed:30± 1cm/min,CRE)	7.50	12.7	kgf

Test Item	Test Method	Test Results
Linting	Average	230.1
	Log 10(lint count)	2.36
Cleanliness-Particulate matter(IPM)	With reference to BS EN 13795-2:2004/ISO 9073-10:2003 method.	2.23

NOTE : 1.The test report merely reflects the test results of the consigned matters of the client and is not a certification of the legitimacy of the related products.

2.The content of this report is invalid if it is not presented as the entire report.

3."IBE":Ignited, But Extinguished.

4.The test was conducted in ISO 17025 accordance laboratory.

- END -

Test Report

REPORT NO: UB/2018/10172

Date: 2018/02/14


Page : 3 of 3



Green Health Canada Inc.

1-1630 Matheson Blvd, Mississauga Ontario, L4W 1Y4 Canada

Sample Photo



This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.